

Art Unit: 3767

EXAMINER'S AMENDMENT

An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Dan Altman on 1/31/2008.

In claim 1, on lines 1 and 2 of the claims, following "the syringe barrel"

--- "being adapted to be" --- has been deleted.

T.S.
3/5/08
In claim 1, on line ²²~~20~~ of the claim, following "deformed when compressed"

--- "." --- has been deleted and

--- "; said cylinder holder comprising: a flange insertion groove for holding the syringe barrel; and a concave portion formed on an inner wall surface of the flange insertion groove to be contacted with the rear surface of the flange of the syringe barrel; whereby, the concave portion is engaged with the projection on the rear surface of the flange when the syringe barrel is mounted in a use position." --- has been inserted.

In claim 9, on line 1 of the claim, following "A cylinder holder"

--- "with a syringe barrel mounted thereon, said cylinder holder" --- has been inserted.

¹⁶
In claim 9, line ~~15~~ of the claim, following "wherein front is considered"

--- "." --- has been deleted and

--- "; said syringe comprising: a flange to be held by the flange insertion groove; and a concave portion formed on the rear surface of the flange, whereby the concave portion is engaged with the

w